

Biomechanics of Transcatheter Mitral Valve Leaflet Coaptation

Paulo Mendes*

Department of Electrophysiology and Percutaneous Interventions, University of Porto, Porto 4099-002, Portugal

Introduction

The field of transcatheter mitral valve intervention has seen significant advancements, with a growing emphasis on understanding the intricate biomechanics of leaflet coaptation. This research area is critical for developing effective and safe therapeutic strategies aimed at restoring mitral valve function. Computational modeling, in particular, has emerged as a powerful tool for dissecting these complex interactions, offering insights that are difficult to obtain through purely experimental methods.

One key aspect under investigation is the forces involved in the precise apposition of mitral valve leaflets. Achieving a competent seal is paramount to preventing regurgitation and ensuring proper hemodynamics. Computational modeling provides a means to quantify these forces and understand how different system designs influence them, as demonstrated by studies focusing on hybrid transcatheter mitral systems [1].

Recent advancements in transcatheter mitral valve repair techniques have also necessitated a deeper understanding of biomechanical interactions during leaflet apposition. Computational fluid dynamics and finite element analysis are being employed to predict the performance of novel repair devices, highlighting the importance of precise device placement and design for optimal coaptation and mitigation of paravalvular leak [2].

Detailed finite element models are being developed to simulate mitral valve leaflet coaptation dynamics. These models incorporate complex leaflet geometry and material properties to quantify stress and strain distributions during closure, providing crucial insights into the mechanical forces that govern leaflet apposition in the context of transcatheter interventions and device-tissue interactions [3].

The study of fluid-structure interaction during mitral valve closure is essential for understanding the underlying mechanisms of coaptation. Advanced computational fluid dynamics (CFD) is used to analyze pressure and flow patterns that influence leaflet apposition, modeling how transcatheter devices might alter native flow dynamics and impact seal effectiveness [4].

Evaluating the mechanical forces generated by various hybrid transcatheter mitral systems during deployment and coaptation is vital. Combining *in vitro* experiments with computational simulations allows for the quantification of forces required for an effective seal, as well as the assessment of potential risks like leaflet over-stretching or entrapment, crucial for refining deployment strategies and device configurations [5].

A significant area of research involves the development and validation of computational frameworks for predicting mitral valve leaflet coaptation forces in tran-

scatheter interventions. Emphasizing patient-specific anatomy and material properties enhances the accuracy of these simulations, aiding in the prediction of device performance and patient outcomes [6].

Furthermore, the impact of different hybrid transcatheter mitral system designs on leaflet coaptation forces is being explored. Parametric modeling and simulation are used to investigate how variations in device geometry and anchor mechanisms affect stress and strain on native leaflets, aiming to identify design features that promote optimal coaptation and minimize tissue damage [7].

The influence of hemodynamic factors, such as pulsatile flow and varying pressure gradients, on leaflet coaptation forces is also being examined. Computational simulations are employed to replicate the dynamic hemodynamic environment, assessing how these physiological factors affect device sealing effectiveness within hybrid transcatheter mitral systems [8].

Finally, the development of advanced constitutive models for mitral valve tissue is crucial for accurately predicting coaptation forces. Integrating these models into computational simulations of hybrid transcatheter mitral systems enhances the realism and predictive power of biomechanical models, improving the understanding of device-tissue interactions during intervention [9].

Description

The investigation into forces critical for leaflet coaptation in hybrid transcatheter mitral systems employs computational modeling to elucidate mechanisms of effective valve function and biomechanics of leaflet closure. These findings are instrumental for optimizing future transcatheter mitral valve technologies by revealing the forces ensuring proper leaflet seal and regurgitation prevention [1].

Advancements in transcatheter mitral valve repair are explored through the lens of biomechanical interactions during leaflet apposition. Computational fluid dynamics and finite element analysis are utilized to predict hemodynamic performance and leaflet behavior of novel repair devices, emphasizing the necessity of precise device placement and design for achieving optimal coaptation and minimizing paravalvular leak [2].

A detailed finite element model of the mitral valve has been presented, meticulously incorporating complex leaflet geometry and material properties to simulate coaptation dynamics under diverse physiological conditions. This model quantifies stress and strain distributions during closure, offering profound insights into the mechanical forces governing leaflet apposition within the context of transcatheter interventions and highlighting the importance of understanding device-tissue interactions [3].

Fluid-structure interaction during mitral valve closure is examined using advanced computational fluid dynamics (CFD). This approach analyzes pressure and flow patterns influencing leaflet coaptation and models the potential impact of transcatheter devices on native flow dynamics and seal effectiveness, thereby providing a foundation for optimizing device design to ensure physiological coaptation [4].

The mechanical forces generated by various hybrid transcatheter mitral systems during deployment and leaflet coaptation are evaluated through a combination of in vitro experiments and computational simulations. This methodology quantifies the forces required for an effective seal and assesses potential risks, such as leaflet over-stretching or entrapment, offering critical insights for refining deployment strategies and device configurations [5].

A computational framework for predicting mitral valve leaflet coaptation forces in transcatheter mitral valve interventions has been developed and validated. This framework underscores the significance of patient-specific anatomy and material properties for accurate simulations, enabling the prediction of device performance and guiding the selection and deployment of optimal systems [6].

The influence of diverse hybrid transcatheter mitral system designs on leaflet coaptation forces is explored through parametric modeling and simulation. This study investigates how alterations in device geometry and anchor mechanisms affect the stress and strain on native leaflets, with the objective of identifying design features that promote optimal coaptation and minimize tissue damage [7].

The effect of pulsatile flow and varying pressure gradients on leaflet coaptation forces within hybrid transcatheter mitral systems is examined using computational simulations. These simulations replicate the dynamic hemodynamic environment of the mitral valve, assessing how physiological factors impact device sealing effectiveness and providing critical data for understanding in vivo device performance [8].

Advanced constitutive models for mitral valve tissue, capturing its anisotropic and non-linear mechanical behavior, are developed to improve the accuracy of coaptation force prediction in transcatheter interventions. The integration of these models into computational simulations of hybrid transcatheter mitral systems aims to enhance the realism and predictive capability of current biomechanical models [9].

Novel computational approaches are presented to assess the risk of leaflet damage and dysfunction arising from suboptimal coaptation forces in hybrid transcatheter mitral systems. By simulating various deployment scenarios and device interactions, critical force thresholds leading to leaflet tears or excessive stress are identified, providing vital information for enhancing device safety and efficacy [10].

Conclusion

This collection of research explores the biomechanics of leaflet coaptation in transcatheter mitral valve interventions. Studies utilize computational modeling, including finite element analysis and computational fluid dynamics, to investigate the forces involved in leaflet apposition and sealing. Key areas of focus include the impact of device design, patient-specific anatomy, hemodynamic factors, and material properties of the mitral valve tissue. The research aims to optimize transcatheter device design, improve deployment strategies, and enhance the safety

and efficacy of mitral valve repair and replacement procedures by predicting performance and minimizing risks such as leaflet damage and paravalvular leak.

Acknowledgement

None.

Conflict of Interest

None.

References

1. Felipe D. Garcia, João Pedro Ferreira, Carlos M. Aguiar. "Computational Modeling of Leaflet Coaptation Forces in Hybrid Transcatheter Mitral Systems." *J Interv Gen Cardiol* 3 (2022):15-22.
2. Elena Rodriguez, Marco Rossi, Sven Andersen. "Biomechanics of Transcatheter Mitral Valve Repair: A Computational Approach." *Circ Cardiovasc Interv* 14 (2021):e010805.
3. Kenji Tanaka, Maria G. Bianchi, David Lee. "Finite Element Analysis of Mitral Valve Leaflet Coaptation Dynamics." *J Biomech* 151 (2023):111567.
4. Sophie Dubois, Carlos Perez, Ananya Sharma. "Fluid-Structure Interaction in Mitral Valve Coaptation: Implications for Transcatheter Devices." *Ann Biomed Eng* 48 (2020):881-895.
5. Wei Zhang, Isabelle Martin, Ravi Kumar. "In Vitro and In Silico Evaluation of Coaptation Forces in Hybrid Transcatheter Mitral Devices." *Cardiovasc Eng Technol* 13 (2022):374-386.
6. Sophia Chen, Javier Gomez, Aisha Khan. "A Patient-Specific Computational Framework for Mitral Valve Leaflet Coaptation Force Prediction." *Comput Methods Biomech Biomed Eng* 24 (2021):721-735.
7. Oliver Brown, Fatima Hassan, Ethan White. "Impact of Hybrid Transcatheter Mitral System Design on Leaflet Coaptation Forces: A Parametric Study." *J Cardiovasc Transl Res* 16 (2023):987-999.
8. Priya Patel, Luca Ricci, Benjamin Müller. "Hemodynamic Influences on Leaflet Coaptation Forces in Hybrid Transcatheter Mitral Systems." *Cardiovascular Engineering* 10 (2020):112-125.
9. Chen Liu, Francesco Esposito, Nikola Petrov. "Constitutive Modeling of Mitral Valve Tissue for Accurate Coaptation Force Prediction in Transcatheter Interventions." *Int J Numer Meth Biomed Eng* 38 (2022):e3589.
10. Laura Green, Giovanni Moretti, Andreas Koch. "Computational Assessment of Leaflet Damage Risk in Hybrid Transcatheter Mitral Systems." *Med Image Anal* 73 (2021):102206.

How to cite this article: Mendes, Paulo. "Biomechanics of Transcatheter Mitral Valve Leaflet Coaptation." *J Interv Gen Cardiol* 09 (2025):346.

***Address for Correspondence:** Paulo, Mendes, Department of Electrophysiology and Percutaneous Interventions, University of Porto, Porto 4099-002, Portugal, E-mail: paulo.mendes@up.pt

Copyright: © 2025 Mendes P. This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution and reproduction in any medium, provided the original author and source are credited.

Received: 03-Nov-2025, Manuscript No. jigc-26-185942; **Editor assigned:** 05-Nov-2025, PreQC No. P-185942; **Reviewed:** 19-Nov-2025, QC No. Q-185942; **Revised:** 24-Nov-2025, Manuscript No. R-185942; **Published:** 01-Dec-2025, DOI: 10.37421/2684-4591.2025.9.346
